



Medical devices — Validation and routine control of ethylene oxide sterilization

TECHNICAL CORRIGENDUM 1

Dispositifs médicaux — Validation et contrôle de routine de la stérilisation à l'oxyde d'éthylène

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 11135:1994 was prepared by Technical Committee ISO/TC 198, Sterilization of health care products.

Page 20

Replace the penultimate equation

$$a = 0,25 \sum_{i=2}^{i=6} (t_{i+1} + t_{i-1})^2 \left(\frac{r_i(n_i - r_i)}{n_i^2(n_i - 1)} \right)$$

by the following:

$$a = 0,25 \sum_{i=2}^{i=6} (t_{i+1} - t_{i-1})^2 \left(\frac{r_i(n_i - r_i)}{n_i^2(n_i - 1)} \right)$$

